



FEBRILE ANTIGENS  
DIRECTIONS FOR USE

**Stained Febrile Antigens: For Widal And Weil-Felix Tests.**

**PRINCIPLE**

The stained antigen suspensions may be used to identify and quantitate specific antibodies in human sera following infection with certain Salmonella, Rickettsiae and Brucellae pathogens. The stained febrile antigens are suitable for both the rapid slide and tube agglutination tests against human sera for the detection of these agglutinins.

**KIT DESCRIPTION**

Lorne Stained Febrile Antigens are for the detection of certain Salmonella, Rickettsiae and Brucellae pathogens. The antigens are suspensions of killed bacteria, stained to enhance the reading of agglutination tests. The blue stained antigens are specific to the somatic "O" antigens and the red stained antigens are specific to the flagellar "H" antigens. Suspensions of Proteus OX2, OX19 and OXK are used to detect rickettsial antibodies. For lot reference number and expiry date see **Vial Labels**.

**STORAGE**

Do not freeze. Reagent vials should be stored at 2 - 8°C on receipt. Prolonged storage at temperatures outside this range may result in accelerated loss of reagent reactivity.

**SPECIMEN COLLECTION**

Specimens should be drawn without anticoagulant using an aseptic phlebotomy technique. Remove serum from clot by centrifugation and store at 2-8°C for 48 hours before performing the test. For longer periods of time the serum must be frozen at or below -20°C. Do not use plasma or heat inactivated, grossly haemolysed, gross lipaemic or contaminated serum specimens.

**PRECAUTIONS**

1. The kit is for *in vitro* diagnostic use only.
2. Do not use kit past expiration date (see **Vial and Box Labels**).
3. The reagents are light sensitive and must be stored in the dark.
4. Do not ingest or inhale aerosols, wash any splashes with copious amounts of water.
5. Protective clothing should be worn when handling the reagents, such as disposable gloves and a laboratory coat.
6. No known tests can guarantee products derived from human or animal sources are free from infectious agents. Care must be taken in the use and disposal of each vial and its contents.

**DISPOSAL OF KIT REAGENT AND DEALING WITH SPILLAGES**

For information on disposal of reagents and decontamination of a spillage site see **Material Safety Data Sheets**, available on request.

**CONTROLS AND ADVICE**

1. It is recommended that known positive and negative controls be tested in parallel with each batch of tests. Tests must be considered invalid if controls do not show expected results.
2. All the reagents must be allowed to reach 18-25°C before use.
3. Shake the reagents well before use to ensure homogeneity.
4. The use of the reagents and the interpretation of results must be carried out by properly trained and qualified personnel in accordance with the requirements of the country where the reagents are in use.
5. The user must determine the suitability of the reagents for use in other techniques.

**MATERIALS AND EQUIPMENT REQUIRED**

- Serological Pipette.
- Small Plastic Test Tubes.
- Agglutination Slides.
- Mixing Sticks.
- 37°C Water Bath.
- 9 g/L saline solution.

**SLIDE TECHNIQUE**

1. Pipette 50 µl of the sample to be tested and 1 drop of each control into separate circles on the slide (when testing for Brucella antibodies it is recommended to reduce the sample volume to 20 µl).
2. Add 50 µl of undiluted antigen suspension to each circle next to the sample to be tested.
3. Mix well using a disposable stirrer and spread the mixture over the entire area enclosed by the circle.
4. Swirl the slide and read the agglutination in each circle macroscopically after one minute.

**SLIDE TITRATION TECHNIQUE**

1. Pipette 80, 40, 20, 10 and 5 µl of undiluted sample to be tested into separate circles on the slide.
2. Pipette 50 µl of the undiluted antigen suspension next to the sample in each circle on the slide.
3. Mix the contents of each circle well using a disposable stirrer and spread the mixture over the entire area enclosed by the circle.
4. Swirl the slide and read the agglutination in each circle macroscopically after one minute.

**INTERPRETATION OF SLIDE TITRATION RESULTS**

1. Agglutination seen in any circle is indicative of the following results should a tube test be carried out:

Volume	80 µl	40 µl	20 µl	10 µl	5 µl
Results	1:20	1:40	1:80	1:160	1:320

2. In this way the slide titration test provides an approximation to the expected results from a corresponding tube test.
3. It is necessary to perform all dilutions in the slide test to obviate the "prozone" effect where higher concentrations of the serum may give a negative result but further dilutions may give a positive result.

**RECOMMENDED TUBE AGGLUTINATION TECHNIQUE**

1. Label 8 small plastic tubes in a rack.
2. Using a pipette dispense 1.9 ml of 9 g/L saline solution into the first tube, and 1.0 ml into the remaining seven.
3. Using a pipette dispense 0.1 ml of the patient's undiluted serum into the first tube.
4. Mix contents well using the pipette, making sure not to create any air bubbles.
5. Dispense 1.0 ml from first tube into second tube and mix well.
6. Dispense 1.0 ml from second tube into third tube and mix well.
7. Continue this method of doubling dilutions up to the seventh tube and then discard 1.0 ml from the seventh tube.
8. The eighth tube will contain only saline as a control and therefore should not contain any serum.
9. Add one drop of the appropriate antigen suspension into each tube and mix well.
10. Incubate the tubes as follows:
  - Somatic "O" antigens and Proteus for 4 hours at 50°C (±2°C).
  - Flagellar "H" antigens for 2 hours at 50°C (±2°C).
  - Brucella Antigen for 24 hours at 37°C (±1°C).
- NB: It is vitally important that when the tubes are placed in a water bath, the level of water should come to approximately 2/3<sup>rd</sup> the way up the level of the tube content. This will maintain convection currents within the tube and thereby obviate false results.
11. Examine the tubes after the appropriate incubation time and check for agglutination.
12. The titre to be taken is the last tube to show agglutination.

**INTERPRETATION OF TUBE RESULTS**

1. Tubes should be read after the recommended incubation time to eliminate the possibility of false results.

2. A somatic reaction (O) is characterised by coarse, compact agglutination, which tends to be difficult to disperse, while flagellar reactions (H) have a characteristic loose, flocculant agglutination.
2. Last tube showing signs of agglutination should be taken as titre for that test. For negative results, all tubes should show no agglutination.
3. Partial or complete agglutination with a variable degree of clearing of the supernatant fluid is recorded as a positive.
4. Titres in excess of 1:80 are usually significant and may reflect recent infection, but low titres can be found in patients.

#### STABILITY OF THE REACTIONS

Slide tests should be interpreted immediately after the 1-minute rotation period to avoid the possibility that a negative result may be incorrectly interpreted as positive due to drying of the reagent.

#### LIMITATIONS

1. It has been found that many serotypes of salmonella possess somatic antigens of the same kind. Agglutination of any of the Salmonella antigens with human serum should not be taken as proof of infection by one particular organism, but rather than as an infection by an organism of a like antigenic structure.
2. A great number of false positive reactions have been reported in healthy individuals when tested with Proteus antigens, especially in slide tests. A titre of less than 1:160 should not be considered significant.
3. False positive or false negative results may also occur due to:
  - Contamination of test materials
  - Improper incubation time or temperature
  - Improper storage of test materials or omission of reagents
  - Deviation from the recommended techniques

#### SPECIFIC PERFORMANCE CHARACTERISTICS

1. The reagents have been characterised by all the procedures mentioned in the **Recommended Techniques**.
2. Prior to release, each lot of Lorne Febrile Antigens is tested by the **Recommended Techniques** to ensure suitable reactivity.
3. There is no International Reference standard for the sensitivity standardisation of these reagents. That's why Lorne uses an internal control that contains animal serum with antibodies of Salmonellas, Brucellas and Proteus that is tested against commercially available reagents of certified potency.

#### DISCLAIMER

1. The user is responsible for the performance of the reagents by any method other than those mentioned in the **Recommended Techniques**.
2. Any deviations should be validated prior to use using established laboratory procedures.

#### BIBLIOGRAPHY

1. Huddleson IF and Bell (1928): J. Infect. Dis. 42: 242.
2. Freter P. (1980). Man. Of Clin. Imm. 2<sup>nd</sup> Ed. ASM Washington DC 460-463.
3. Weil E and Felix A. Wein Klin. 29: 974 (1916).
4. Cruickshank R (1965). Med. Mic. 11<sup>th</sup> Ed. 907.

#### AVAILABLE REAGENT AND KIT SIZES

##### A. Febrile Antigens:

	Volume	Catalogue Number
Salmonella Typhi H	5 ml	502005A
Salmonella paratyphi AH	5 ml	504005A
Salmonella paratyphi BH	5 ml	506005A
Salmonella paratyphi CH	5 ml	508005A
Salmonella Typhi O	5 ml	510005A
Salmonella paratyphi AO	5 ml	512005A
Salmonella paratyphi BO	5 ml	514005A
Salmonella paratyphi CO	5 ml	516005A
Brucella abortus	5 ml	518005A
Brucella melitensis	5 ml	520005A
Proteus OX2	5 ml	522005A
Proteus OX19	5 ml	524005A
Proteus OXK	5 ml	526005A

##### B. Febrile Antigen Kits:

	Kit Size	Catalogue Number
Febrile Bacterial Kit	8 x 5 ml	532040A
Febrile Bacterial Kit + Controls	8 x 5 ml + 2 x 1ml	532042A

##### C. Febrile Antigen Controls:

	Kit Size	Catalogue Number
Febrile Positive Control	1 ml	536001A
Febrile Negative Control	1 ml	537001A








All stained febrile antigens are available in bulk quantities of 500 ml or 1 litre to special order.

For the availability of other sizes, please contact:

#### Lorne Laboratories Limited

Unit 1 Danehill  
Cutbush Park Industrial Estate  
Lower Earley, Reading,  
Berkshire, RG6 4UT  
England  
Tel: +44 (0) 118 921 2264  
Fax: +44 (0) 118 986 4518  
E-mail: [info@lornelabs.com](mailto:info@lornelabs.com)

#### TABLE OF SYMBOLS

	Batch Number		<i>in-vitro</i> Diagnostic
	Catalogue Reference		Store At
	Expiry Date		Manufacturer
	Read Pack Insert		