

MONOCLONAL BLOOD GROUPING REAGENTS  
DIRECTIONS FOR USE

**Anti-Le<sup>b</sup> Monoclonal: For Tube Technique.**

**SUMMARY**

The Lewis system antigens are not an integral part of the red cell membrane and are produced by tissue cells and found primarily in plasma and watery secretions. Red cells acquire Lewis antigens by adsorption from surrounding plasma. The amount of Lewis antigen expressed on a cell can vary with the cell's ABO phenotype. Anti-Le<sup>a</sup> and Anti-Le<sup>b</sup> have not been associated with Haemolytic Disease of the Newborn.

Anti-Le <sup>a</sup>	Anti-Le <sup>b</sup>	Phenotype	Caucasians %	Afro-Americans %
+	0	Le(a+b-)	22	23
0	+	Le(a-b+)	72	55
0	0	Le(a-b)	6	22
+	+	Le(a+b+)	Rare	Rare

**PRINCIPLE**

The reagent will cause agglutination (clumping) of test red cells, that carry the Lewis b antigen, after centrifugation. No agglutination generally indicates the absence of the corresponding Lewis antigen (see **Limitations**).

**REAGENTS**

Lorne Monoclonal Anti-Le<sup>b</sup> blood grouping reagent contains mouse monoclonal IgM antibodies, diluted in a phosphate buffer containing sodium chloride, EDTA, bovine albumin and macromolecular potentiators. Anti-Le<sup>b</sup> is manufactured with Clone LEB2. The reagent is supplied at optimal dilution for use with all the recommended techniques stated below without the need for further dilution or addition. For lot reference number and expiry date see **Vial Label**.

**STORAGE**

Reagent vials should be stored at 2 - 8°C on receipt. Prolonged storage at temperatures outside this range may result in accelerated loss of reagent reactivity. This reagent has undergone transportation stability studies at 37°C and -25°C as described in document EN13640:2002.

**SAMPLE COLLECTION AND PREPARATION**

Blood samples drawn with or without anticoagulant may be used for antigen typing. If testing is delayed then store specimens at 2-8°C. All blood specimens should be typed as soon as possible after collection. Donor blood collected in ACD, CPD, CPDA-1, CP2D, CP2D with AS-3, CPD with AS-1, and CPD with AS-5 may be tested until the expiration date of the donation. Specimens displaying gross haemolysis or microbial contamination should not be tested with this reagent.

**PRECAUTIONS**

1. The reagent is intended for *in vitro* diagnostic use only.
2. If a reagent vial is cracked or leaking, discard the contents immediately.
3. Do not use the reagents past the expiration date (see **Vial Label**).
4. Do not use the reagents if a precipitate is present.
5. Protective clothing should be worn when handling the reagents, such as disposable gloves and a laboratory coat.
6. The reagent has been filtered through a 0.2 µm capsule to reduce the bio-burden. Once a vial has been opened the contents should remain viable up until the expiry date as long as there is no marked turbidity, which can indicate reagent deterioration or contamination.
7. The reagent contains <0.1% sodium azide. Sodium azide may be toxic if ingested and may react with lead and copper plumbing to form explosive metal azides. On disposal flush away with large volumes of water.
8. No known tests can guarantee that products derived from human or animal sources are free from infectious agents. Care must be taken in the use and disposal of each vial and its contents.

**DISPOSAL OF REAGENT AND DEALING WITH SPILLAGES**

For information on disposal of the reagent and decontamination of a spillage site see **Material Safety Data Sheets**, available on request.

**CONTROLS AND ADVICE**

1. It is recommended a positive and negative control be tested in parallel with each batch of tests. Tests must be considered invalid if controls do not show expected results.
2. In the **Tube Technique** one volume is approximately 50µl when using the vial dropper provided.
3. The use of the reagent and the interpretation of results must be carried out by properly trained and qualified personnel in accordance with the requirements of the country where the reagent is in use.
4. The user must determine suitability of the reagent for use in other techniques.

**REAGENTS AND MATERIALS REQUIRED**

- Glass test tubes (10 x 75 mm or 12 x 75 mm).
- PBS solution (pH 6.8-7.2) or Isotonic saline solution (pH 6.5-7.5).
- Positive and negative control red cells:  
Le(b+)(positive control) and Le(b-) (negative control).
- Test tube centrifuge.
- Volumetric pipettes.

**RECOMMENDED TECHNIQUE**

**A. Tube Technique**

1. Prepare a 2-5% suspension of washed red cells in PBS or Isotonic saline.
2. Place in a labelled test tube: 1 volume of Lorne Anti-Le<sup>b</sup> reagent and 1 volume of red cell suspension.
3. Mix thoroughly and incubate at room temperature for 15 minutes.
4. Centrifuge all tubes for 20 seconds at 1000 rcf or for a suitable alternative time and force.
5. Gently resuspend red cell button and read macroscopically for agglutination

**INTERPRETATION OF TEST RESULTS**

1. **Positive:** Agglutination of the test red cells constitutes a positive test result and within accepted limitations of test procedure, indicates the presence of the Lewis b antigen on the test red cells.
2. **Negative:** No agglutination of the test red cells constitutes a negative result and within accepted limitations of the test procedure, indicates absence of the Lewis b antigen on the test red cells.

**STABILITY OF THE REACTIONS**

1. Tests should be read immediately after centrifugation. Delays may result in dissociation of antigen-antibody complexes leading to false negative, or weak positive reactions.
2. Caution should be exercised in the interpretation of results of tests performed at temperatures other than those recommended.

**LIMITATIONS**

1. Lorne Anti-Le<sup>b</sup> reagent must only be used with washed red cells suspended in PBS or Isotonic saline because Lewis antigens are present in plasma. Cells suspended in plasma/serum **cannot** be used since the soluble antigen present may neutralise the test reagent, giving false negative results.
2. Weaker reactions may occur when Anti-Le<sup>b</sup> is tested against A<sub>1</sub> or A<sub>1</sub>B Le(b+) red cells because amount of Lewis antigen expressed on red cell can vary with cell's ABO phenotype.
3. Red cells of most new-borns will type Le(a-b-) with monoclonal or human anti-Lewis reagents.
4. The Lewis phenotypes of children under six years of age cannot be accurately determined. Red cell Lewis antigens are weaker during pregnancy and some women with red cells of the Le(a-b+) phenotype may type as Le(a-b-) whilst pregnant.
5. Stored blood may give weaker reactions than fresh blood
6. False positive or false negative results may also occur due to:
  - Contamination of test materials
  - Improper storage, cell concentration, incubation time or temperature
  - Improper or excessive centrifugation
  - Deviation from the recommended techniques

**SPECIFIC PERFORMANCE CHARACTERISTICS**

1. The reagents have been characterised by all the procedures mentioned in the **Recommended Techniques**.
2. Prior to release, each lot of Lorne Monoclonal Anti-Le<sup>b</sup> is tested by the **Recommended Technique** against a panel of antigen-positive red cells to ensure suitable reactivity.
3. Specificity of source monoclonal antibodies is demonstrated using a panel of antigen-negative cells.
4. The Quality Control of the reagents was performed using red cells that had been washed twice with PBS or Isotonic saline prior to use.
5. The reagents comply with the recommendations contained in the latest issue of the Guidelines for the UK Blood Transfusion Services.

**DISCLAIMER**

1. The user is responsible for the performance of the reagent by any method other than those mentioned in the **Recommended Technique**.
2. Any deviations from the **Recommended Technique** should be validated prior to use<sup>6</sup>.

**BIBLIOGRAPHY**

1. Kholer G, Milstein C. Continuous culture of fused cells secreting antibody of predefined specificity. Nature 1975; **256**, 495-497

2. Mollison PL. Blood Transfusion in Clinical Medicine, 8<sup>th</sup> Edition, Blackwell Scientific, Oxford 1987; Chapter 7.
3. Issitt PD. Applied Blood Group Serology, 3<sup>rd</sup> Edition. Montgomery Scientific, Miami 1985; Chapter 6
4. BSCH Blood Transfusion Task Force. Guidelines for microplate techniques in liquid-phase blood grouping and antibody screening, Clinical Laboratory Haematology 1990; **12**, 437-460.
5. Guidelines for the Blood Transfusion Service in the United Kingdom. H.M.S.O. Current Edition.
6. British Committee for Standards in Haematology, Blood Transfusion Task Force. Recommendations for evaluation, validation and implementation of new techniques for blood grouping, antibody screening and cross matching. Transfusion Medicine, 1995, **5**, 145-150.

**AVAILABLE REAGENT SIZES**


	Vial Size	Catalogue Number
Anti-Le <sup>b</sup> Monoclonal	2 ml	631002
	1000 ml	631000*

\*This size is For Further Manufacturing Use (FFMU) only is therefore not CE marked.

For the availability of other sizes, please contact:

**Lorne Laboratories Limited**  
 Unit 1 Cutbush Park Industrial Estate  
 Danehill  
 Lower Earley, Reading,  
 Berkshire, RG6 4UT  
 England  
 Tel: +44 (0) 118 921 2264  
 Fax: +44 (0) 118 986 4518  
 E-mail: info@lornelabs.com

**TABLE OF SYMBOLS**

	Batch Number		<i>in-vitro</i> Diagnostic
	Catalogue Reference		Store At
	Expiry Date		Manufacturer
	Read Pack Insert		