



EC CERTIFICATE

Lorne Laboratories Ltd

Unit 1 Cutbush Park Industrial Estate,
Danehill, Lower Earley, Berkshire RG6 4UT, UK

EC Certificate - Full Quality Assurance System Approval Certificate

Annex IV, (excluding sections 4 and 6) of Council Directive 98/79/EC on In Vitro
Diagnostic Medical Devices

Scope of Certificate:

**The design and manufacture of in vitro diagnostic reagents for
identification of blood groups**

Device Classification:

Annex II, List A and B

Device Descriptions:

Please refer to Attachment 1

Model:

Please refer to Attachment 1

File Number	A12241	Cycle Start Date	23 May 2017
Certificate No.	354.170425	Effective Date	23 May 2017
		Expiry Date	22 May 2022

Authorised by

B. Rodgers

Certification Manager

For and on Behalf of UL International (UK) Ltd

We hereby declare that an examination of the full quality assurance system has been carried out per report 11640248, following the requirements of the national legislation to which the undersigned is subject, transposing Annex IV (with the exemption of sections 4 and 6) of Council Directive 98/79/EC on In Vitro Diagnostic Medical Devices. We certify that the full quality assurance system conforms with the relevant provisions of the aforementioned directive and is subject to periodic surveillance as required by 98/79/EC, Annex IV, Section 5. For Annex II, List A devices where they are covered by this certificate, an EC Design Examination certificate according to 98/79/EC, Annex IV, Section 4 is required. This certificate is issued with 1 attachment listing model numbers.

Notified Body

0843

UL International (UK) Limited
Wonersh House, The Guildway, Old Portsmouth Road,
Guildford, Surrey, GU3 1LR, United Kingdom



EC CERTIFICATE

Lorne Laboratories Ltd

Unit 1 Cutbush Park Industrial Estate,
Danehill, Lower Earley, Berkshire RG6 4UT, UK

Attachment 1 of 1

The products detailed below are covered under the scope of this certificate

Device Description	Model	Classification
Anti-A Monoclonal	600005/600010/600000	Annex II List A
Anti-B Monoclonal	610005/610010/610000	Annex II List A
Anti-A,B Monoclonal	620005/620010/620000	Annex II List A
Anti-C Monoclonal	690005	Annex II List A
Anti-E Monoclonal	691005	Annex II List A
Anti-c Monoclonal	692005	Annex II List A
Anti-e Monoclonal	693005	Annex II List A
Anti-K Monoclonal	760005/760010	Annex II List A
Anti-D Clone 2 Monoclonal	710010/710000	Annex II List A
Anti-D Clone 1 Monoclonal	730010/730000	Annex II List A
Anti-D Duoclone Monoclonal	740010/740000	Annex II List A
Anti-Jka Polyclonal	323002/323000	Annex II List B
Anti-Jkb Polyclonal	324002/324000	Annex II List B
Anti-Fyb Polyclonal	317002/317000	Annex II List B
AHG Elite Clear	415010/415100/415000	Annex II List B
AHG Elite Green	435010/435100/435000	Annex II List B
Anti-Fya Monoclonal	774000/774002	Annex II List B
Anti-C+D+E Monoclonal	700005/700010/700000	Annex II List A
Anti-Human IgG Clear	401010/401000	Annex II List B
Anti-Human IgG Green	402010/402000	Annex II List B
Monoclonal Rh Control	640010	Annex II List A
Monoclonal D Negative Control	650010	Annex II List A

File Number A12241
Certificate No. 354.170425

Cycle Start Date 23 May 2017
Effective Date 23 May 2017
Expiry Date 22 May 2022

Authorised by

B. Rodgers
Certification Manager
For and on Behalf of UL International (UK) Ltd

Notified Body

0843